

# Cleveland Open Cup Flash Point Tester *Model: Flash-92m*



Acute Instruments Pvt Ltd  
 31, 33 & 35 Ground Floor Sai Chambers,  
 Plot 44, Sector 11, CBD Belapur, Navi Mumbai, 400614, Maharashtra, India  
 Email: sales@acuteinstruments.com  
 Tel: 022 49629780 / 022 49709780  
 www.acuteinstruments.co.in



<b>Methods</b>	ASTM D 92, ISO2592, IS1448(P: 69) & IP36
<b>Cooling</b>	Axial Fan @230VAC
<b>Heating</b>	Coil Heater 1000 watts @ 230 Volts
<b>Heat Rate</b>	Heat rate control with precise ten turn potentiometer
<b>Cooling</b>	Axial Fan @230VAC
<b>Ignition</b>	Gas Igniter - Precise control / Motorised rotation
<b>Test</b>	Manual dipping, heater can be switched off by using toggle switch.
<b>Lid Assembly</b>	Permanently fixed sample temperature sensor and Igniter for easy operation
<b>Sample Temperature</b>	Glass PT 100 , Class A, with digital indication / Thermometer interchangeable
<b>Flash Detection</b>	Manual / Flash Point reading lock by push button
<b>Dimensions</b>	W: 220 mm X D: 300 mm X H: 400 mm
<b>Power Supply</b>	230Volts 50 Hz 1200Watts