

Pensky Martens Closed Cup Flash Point Tester Model: Flash-93m



Acute Instruments Pvt Ltd
 31, 33 & 35 Ground Floor Sai Chambers,
 Plot 44, Sector 11, CBD Belapur, Navi Mumbai, 400614, Maharashtra, India
 Email: sales@acuteinstruments.com
 Tel: 022 49629780 / 022 49709780
 www.acuteinstruments.co.in



Methods	ASTM D93-A,B&C; ISO2719,IP34, IS-1448(P-21)
Cooling	Axial Fan @230VAC
Heating	Band Heater 350 watts @ 230 VAC
Heat Rate	Heat rate control with precise ten turn potentiometer
Stirring	Speed Control by Precise Potentiometer from 0 to 300 rpm,
Ignition	Gas Igniter - Precise control for Gas igniter
Test	Manual dipping, heater and stirrer can be switched off by using toggle switch.
Lid Assembly	Permanently fixed sample cup cover with Stirrer, Thermometer / sample temperature sensor and Igniter for easy operation
Sample Temperature	SS PT 100, Class A, with digital indication / Thermometer interchangeable
Stirrer RPM	Digital Indication of RPM
Dimensions	W: 220 mm X D: 300 mm X H: 325 mm
Power Supply	230 Volts 50 Hz 600Watts